



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/691,859	10/23/2003	Charmaine K. Harris	P0011285.00	6734

27581 7590 06/04/2007
MEDTRONIC, INC.
710 MEDTRONIC PARKWAY NE
MINNEAPOLIS, MN 55432-9924

EXAMINER

ALTER, ALYSSA M

ART UNIT	PAPER NUMBER
----------	--------------

3762

MAIL DATE	DELIVERY MODE
-----------	---------------

06/04/2007

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary

Application No.

10/691,859

Applicant(s)

HARRIS, CHARMAINE K.

Examiner

Alyssa M. Alter

Art Unit

3762

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 17 August 2006.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-29 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-29 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 23 October 2003 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Response to Arguments

Applicant's arguments with respect to claims 1-27 have been considered but are moot in view of the new ground(s) of rejection necessitated by amendment.

Claim Objections

Claims 24-27 are objected to because of the following informalities: claims numbers 24 and 25 repeated and thus cause an incorrect numbering sequence. For examination purposes, the examiner has considered the second "24" to be --26-- , the second "25" to be --27--, "26" to be --28-- and "27" to be --29--. Appropriate correction is required.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

1. Claims 1-2, 5-8, 10-15, 17-20, 22-26 and 28-29 are rejected under 35

U.S.C. 102(e) as being anticipated by King (US Patent Publication 20030088301 A1).

King discloses a paddle electrode with wire like extendable members, as seen in figure

Art Unit: 3762

5. The examiner considers the electrode lead to have an imaginary centerline at the midline.

Furthermore, King discloses on page 5, paragraph 58, "the implanting physician could selectively pass each member out the tip of the lead 14, and would use fluoroscopy to determine the direction of curving and lateral motion, using rotational torque on each member to optimize its position". Additionally, "since the member's tip 35 is metallic, it will be easily visible on fluoroscopy" (page 6, paragraph 63). Therefore, King discloses an orientation marker.

As to claims 3 and 15, as depicted in figure 3, the lead has extendable members 32, with electrodes 33. Since the electrodes are extendable, they are "recessed relative to the major surface".

As to claims 5-8 and 17-20, "the part of an extendable member that curls into a 2-dimensional conductive pad or 3-dimensional electrode is composed of a robust and safe material, such as platinum or platinum/iridium"(page 3, paragraph 22).

As to claim 10-12 and 22-24, "each member may have an asymmetry to match an asymmetry in its channel, so that its deployment is in a fixed direction from the lead body " (page 3, paragraph 23).

The applied reference has a common assignee with the instant application. Based upon the earlier effective U.S. filing date of the reference, it constitutes prior art under 35 U.S.C. 102(e). This rejection under 35 U.S.C. 102(e) might be overcome either by a showing under 37 CFR 1.132 that any invention disclosed but not claimed in

Art Unit: 3762

the reference was derived from the inventor of this application and is thus not the invention "by another," or by an appropriate showing under 37 CFR 1.131.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

1. Claims 4, 16 and 26 are rejected under 35 U.S.C. 102(e) as anticipated by or, in the alternative, under 35 U.S.C. 103(a) as obvious over King (US Patent Publication 20030088301 A1). Since King discloses an orientation marker, this marker can be used to identify the model or serial number of the lead.

In the alternative, although King could employ the orientation maker to determine the model or serial number, King does not disclose the serial number or model number as the orientation marker. It would have been obvious to one having ordinary skill in the art at the time the invention was made to modify the lead as taught by King with a lead identification information since it was known in the art that the retrieval of lead information is important for determining system compliance and manufacturing defects.

2. Claims 9 and 21 are rejected under 35 U.S.C. 103(a) as being unpatentable over King (US Patent Publication 20030088301 A1). King discloses the claimed invention except for transparent polyurethane material. It would have been obvious to one having ordinary skill in the art at the time the invention was made to modify the materials, since

Art Unit: 3762

it has been held to be within the general skill of a worker in the art to select a known material on the basis of its suitability for the intended use as a matter of obvious design choice. *In re Leshin*, 125 USPQ 416 (See MPEP 2144.07). Additionally, it is well known in the art because the transparent polyurethane material is biocompatible and enables the observation of defects in the lead.

Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than **SIX MONTHS** from the date of this final action.

Art Unit: 3762

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Alyssa M. Alter whose telephone number is (571) 272-4939. The examiner can normally be reached on M-F 9am to 4pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Angela Sykes can be reached on (571) 272-4955. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.



Alyssa M Alter
Examiner
Art Unit 3762

GEORGE R. EVANISKO
PRIMARY EXAMINER

5/29/17